



**GOVERNMENT OF GOA
Medical Store Depot (MSD),**

Directorate of Health Services, Campal, Panaji, Goa 403 001

Email: msddhsgoa@yahoo.in

Ph. Nos. 2225646/5540/5668

No. 92/DHS/MSD/PT/2021-22/1053

Dated:15/09/2021

CORRIGENDUM

Ref: No. : Tender Notice no. 92/DHS/MSD/PT/2021-22/852 Dated: 11/08/2021

In the Tender Notice No 92/DHS/MSD/PT/2021-22/852 Dated: 11 /08 /2021, following changes in the specification of Annexure A (Ophthalmic Equipments) are amended which may be read as follows:

Page No	Existing Specifications	Amended to
Annexure A (Ophthalmic Equipment) Sr. No. 5 Flash autoclave		
18	<ul style="list-style-type: none">• Simple and user friendly- Class B• Should allow effective sterilization of hollow tubings• Wrapped or pouched instruments• Safe and long lasting chamber SS304• LCD control panel• Lid lock safety• Electrical safety by MCB• Automatic water quality check• Overpressure safety• Capacity 22 L• Frontloading type with 3 Trays• Inner dimension 250x450 mm• Temperature range upto 134 degree Celsius• Operating pressure 30/206.8 psi/Kpa	<ul style="list-style-type: none">• Simple and user friendly- Class B - pre vacuum and post vacuum cycles.• Should allow effective sterilization of hollow tubings• Wrapped or pouched instruments• Safe and long lasting chamber SS304• LCD control panel displaying data like temperature, pressure, cycle, time etc.• Lid lock safety• Electrical safety by MCB• Automatic water quality check should have sensor for detection of poor input water quality.• Overpressure safety• Capacity 22 L• Frontloading type with 3 Trays• Inner dimension 250x450 mm• Temperature range upto 134 degree Celsius• Operating pressure 30/206.8 psi/Kpa

Sd/-
Director of Health Services